


Linear Roughness Evaluation as per ASME B46.1



Date:		Units: microns
Laboratory:	Clemex Technologies	Magnification: 100x
Department:	Quality Control	Calibration: 1.2676 $\mu\text{m}/\text{pixel}$
User:		# of fields: 1
Sample ID:	2 sided coating	Total area: 493616.32

	Rq	Ra	Rt	Rp	Rv	L	N
Top	7.18	5.47	36.63	15.28	-21.35	808.73	320
Bottom	9.39	7.29	48.35	21.43	-26.92	808.73	320

Top edge: Distance from the mean line in default units



Bottom edge: Distance from the mean line in default units



Comments: